## History of Measurement Systems with Applications

Jorge Luis Romeu, Ph.D. Dept. MAE, Syracuse U. SUNY Emeritus Faculty Presented at: ASQ Binghamton Conference March 27, 2020

### Borrowed from the Presentation: Measurement & Traceability

Minnesota Department of Commerce Weights and Measures Division Carol Hockert, Director With Acknowledgements to: Georgia Harris, National Institute of Standards and Technology Elizabeth Gentry, Oklahoma Bureau of Standards

Historical References
Biblical
U.S. Constitution
Current Authority in State Law
MN Statute Chapter 239
MN Rules Chapter 7601

## For Commerce to Extend Beyond Barter



Currency one can trust and



Measurements one can trust

are required

## Weights & Measures

Ensures equity in the marketplace Impacts >\$25 billion in Commerce ~70,000 devices tested annually Metrology Lab Calibrates weights, volume measures, thermometers, tape measures Petroleum Quality Lab Samples from terminals, gas stations, complaints

Test and Inspect Scales Grain elevators Livestock Railway, vehicle Grocery, hardware, jewelry, more Test and Inspect Meters Gasoline, diesel Milk, ethanol, biodiesel Jet fuel, kerosene, fuel oil, chemical

Package Inspections Labeling requirements Accurate tare taken on packages Selling according to correct unit of measure Grocery, hardware, lumber, any commodity sold by measure or count Interesting investigations Oysters (out of shell) packed in fluid Milk

Metrology Laboratory Commercial calibrations – field inspectors, agents Calibration for industry ~20,000 calibrations annually Petroleum Laboratory ASTM tests Analytical laboratory equipment Pull bad product off-sale

### Measurement

### Lord Kelvin said:

"When you can measure what you are speaking about, and express it in numbers, you know something about it; but when you cannot measure it, when you cannot express it in numbers, your knowledge is of a meagre and unsatisfactory kind..."

## History of Measurement

Egyptian Cubit Dependent on the Pharaoh Traceable Recalibration required Used to build Pyramids Accuracy of 0.05% Historically based on physical artifacts Now based on intrinsic standards Exception?

## Mass Standards

Based on the platinum-iridium cylinder locked up in a vault at the International Bureau of Weights and Measures (BIPM) Current research conducted Watt Balance Counting atoms (silicon) Neither are as accurate as the current standard Most still rely on calibrated pieces of metal

## Metrology

Study or science of measurement Measurements: Associated Uncertainties Traceable Affect all aspects of life Important to engineers...no matter what type of job you have. SI base units

### Measurements

What are the consequences of a bad measurement? Commerce Health and Medicine Manufacturing Technology Aerospace, DOD, NASA, DOE

### Good measurements

What accuracy do you need?
Parts per million
Parts per hundred
How precise is your measurement?
Repeatability
Reproducibility

Traceability

## Traceability

International Vocabulary of Metrology (VIM) Definition:

"Property of the result of measurement or the value of a standard whereby it can be related to stated references, usually national or international standards, through an unbroken chain of comparisons all having stated uncertainties."

## Criteria for Traceability<sup>1</sup>

Result of a measurement or the value of a standard

- 1. Unbroken chain of comparisons
- 2. Stated Uncertainties
- 3. Documented
- 4. Competence
- 5. Reference to national or international standards or SI Units
- 6. Periodic Verification
- 7. Measurement Assurance

<sup>1</sup>NIST IR 6969, GMP 13, based on ILAC G2 with additional requirements

### Traceability – Why do we need it?

Laws and contracts require it
Customers ask for it
Standards specify it
Quality measurements ensure it

### Measurement Traceability for Weights & Measures



Uncertainty FormulaCombined uncertainty

$$u_c = \sqrt{u_s^2 + s_p^2 + u_d^2 + u_o^2}$$

• Expanded uncertainty (what is reported) 95% or 99% confidence interval  $U = u_c \times k$ 

where k = 2 or k = 3

## Competence

Measuring devices should be used by trained personnel Procedures Understanding limits, uncertainty Understanding what accuracy is required Documented training Personnel records Training file Interdependent on others

## **Periodic Verification**

Calibrate and then Recalibrate at specified intervals

- Dependent on time and use
  Set by the user of the device or by law
- Commerce requires annual verification
- Volumetric glassware may have 10 year cycle
- Thermocouples may have 3 month cycle
- Consequences of long intervals

### **Measurement Assurance**

Quality Control
 Check standards
 Interlaboratory comparisons (Round robins)

Control charts
 Standards
 Processes

### Interlaboratory<sub>c</sub> Comparisons



LAB/Operator

## Summary

Measurements are important Historically, today, tomorrow... Traceability is fundamental to making good measurements Link to SI units Stated Uncertainties Documented Procedures/Training Periodic Verification Measurement Assurance

From here on, Romeu's Students and Coursework Material

## Some Examples of Gage R&R

Some Definitions
Quantitative Analysis
Qualitative Analisis
By Minitab Stats SW

### Fig. 13-7



Fig. 15-2a



Fig. 15-2b



Fig. 15-2c



Fig. 15-2d



### Fig. 15-2e





Fig. 15-3

### Validate Measurement System Attribute Data Analysis – MSA 1

Sample #	Expert	Oper	ator 1	Op	perator 2	Oper	ator 3	
		Try 1	Try 2	Try 1	Try 2	Try 1	Try 2	
1	Blister	Blister	Blister	Blister	Blister	Blister	Blister	
2	Good	Light Ed.	Good	Good	Good	Good	Good	R.
3	Drip	Drip	Drip 🗧	Dirt	Dirt	Drip	Drip	
4	Dirt	Contam	Dirt	Dirt	Dirt	Dirt	Dirt /	
5	Contam	Contam	Contam (	Good	Good	Over Run	Over Run /	
6	Blister	Blister	Blister /	Blister	Blister	Dirt	Blister	
7	Good	Good	Good /	Good	Good	Good	Good /	
8	Dirt	Light Ed.	Contam	Good	Good	Dirt	Dirt /	
9	Good	Good	Drip /	Dirt	Good	Drip	Drip	
10	Good	Orange P.	Good	Good	Good	Good	Good	
11	Dirt	Dirt	Good	Dirt	Dirt	Dirt	Dirt	
12	Good	Contam	Light Ed.	Good	Good	Over Run	Over Run	
13	Good	Light Ed.	Light Ed.	Good	Good	Light Ed.	Over Run	
14	Contam	Contam	Contam	Good	Good	Good	Good	
15	Drip	Drip	Light Ed.	Dirt	Good	Drip	Drip	
16	Light Ed.	Light Ed.	Light Ed.	Good	Good	Good	Good	
17	Dirt	Contam	Contam	Good	Dirt	Dirt	Dirt	
18	Dirt	Contam /	Contam	Dirt	Dirt	Dirt	Dirt 众	
19	Blister	Blister	Good	Good	Blister	Blister	Blister	
20	Good	Good	Good	Good	Good	Orange P.	Orange P.	

#### Fig. 15-9

### **Attribute Data Analysis-MSA 1 Results**

#### **Within Appraiser**

Assessment Agreement

Appraiser #	# Inspected #	Matched	Percent (%)	95.0% CI
1	20	11	55.0	(31.5, 76.9)
2	20	16	80.0	(56.3, 94.3)
3	20	18	90.0	( 68.3, 98.8)

# Matched: Appraiser agrees with him/herself across trials. **Each Appraiser vs Standard** 

Assessment Agreement

Appraiser #	Inspected #	Matched	Percent (%)
1	20	8	40.0
2	20	11	55.0
3	20	12	60.0

#### **All Appraisers vs Standard**

Assessment Agreement

# Inspected	# Matched	Percent (%)	95.0% CI
20	2	10.0	( 1.2, /
31.7)			

# Matched: All appraisers' assessments agree with standard

95.0% CI Note: 38% were called bad that were good.
(19.1, 63.22% were called good that were bad. This
(31.5, 76 potentially could yield an improvement in the
(36.1, 80.9 efect rate by 16%

# Matched: Appraiser's assessment across trials agrees with standard. **Between Appraisers** 

Assessment Agreement

 # Inspected
 # Matched
 Percent (%)
 95.0% CI

 20
 2
 10.0
 (1.2, 31.7)

# Matched: All appraisers' assessments agree with each other.

### **Attribute Data Analysis-MSA 2 Results**

#### Within Appraiser

Assessment Aareement

#### All Appraisers vs Standard

	.g. een en e				Assessment A	Agreement		
Appraiser # 1 1 2 3	Inspected 20 20 20 20	# Matched 16 19 20	Percent (%) 80.0 95.0 100.0	95.0% CI (56.3,94.3) (75.1,99.9) (86.1,100.0)	# Inspected 20	# Matched 13	Percent (%) 65.0	95.0% CI ( 40.8, 84.6)
					# Matcheu. P		assessments a	iuree with standar

# Matched: Appraiser agrees with him/herself across trials.

#### **Each Appraiser vs Standard**

Assessment Agreement

Appraiser	# Inspected	# Matched	Percent (%)	95.0% CI
1	20	15	75.0	( 50.9, /91.3
2	20	19	95.0	(75.1, 99.9
3	20	18	90.0	( 68.3, 98.8

# Matched: Appraiser's assessment across trials agrees with standard.

#### **Between Appraisers**

Assessment Agreement

# Inspected # Matched Percent (%) 95.0% CI 20 13 65.0 (40.8, 84.6)

Continue Q.C. training of inspectors to bring up agreement between all appraisers and standard.

**Conclusion: MSA is greatly improved over first.** 

# Matched: All appraisers' assessments agree with each other.

# Student Group Gage R&R Analysis Example

Group 1 Joi Vis Jai Ale Hu Me

Jorge Vishnu James Alekhya Hui Chen Meng An

### Gage R&R Considerations

- Number of Operators: 3 (Officials)
- Number of Procedures: 10
- Number of Trials: 2
- Number of Samples: 20

Refugees claiming for entry need to submit migration application. There are 3 officials handling the application documents and 10 procedures about filling in the files in different orders. We took 2 samples for different officials and procedures relatively to do the Gage R&R analysis.

### Appendix I.- Gage R&R Data

	Procedure	Official	Time (hours)
1	1	1	0.65
2	1	1	0.60
3	2	1	1.00
4	2	1	1.00
5	3	1	0.85
6	3	1	0.80
7	4	1	0.85
8	4	1	0.95
9	5	1	0.55
10	5	1	0.45
11	6	.1	1.00
12	6	1	1.00
13	7	1	0.95
14	7	1	0.95
15	8	1	0.85
16	8	1	0.80
17	9	1	1.00
18	9	1	1.00
19	10	1	0.60
20	10	1	0.70
21	1	2	0.55
22	1	2	0.55
23	2	2	1.05
24	2	2	0.95
25	3	2	0.80
26	3	2	0.75
27	4	2	0.80
28	4	2	0.75
29	5	2	0.40
30	5	2	0.40

	Procedure	Official	Time (hours)
31	6	2	1.00
32	6	2	1.05
33	7	2	0.95
34	7'	2	0.90
35	8	2	0.75
36	8	2	0.70
37	9	2	1.00
38	9	2	0.95
39	10	2	0.55
40	10	2	0.50
41	1	3	0.50
42	1	3	0.55
43	2	3	1.05
44	2	3	1.00
45	3	3	0.80
46	3	3	0.80
47	4	3	0.80
48	4	3	0.80
49	5	3	0.45
50	5	3	0.50
51	6	3	1.00
52	6	3	1.05
53	7	3	0.95
54	7	3	0.95
55	8	3	0.80
56	8	3	0.80
57	9	3	1.05
58	9	3	1.05
59	10	3	0.85
60	10	3	0.80



#### **Results for; Worksheet 2**

#### Gage R&R Study - XBar/R Method

		<pre>%Contribution</pre>
Source	VarComp	(of VarComp)
Total Gage R&R	0.0020816	6.32
Repeatability	0.0011541	3.51
Reproducibility	0.0009275	2.82
Part-To-Part	0.0308455	93.68
Total Variation	0.0329271	100.00

		Study Var	%Study Var
Source	StdDev (SD)	(6 × SD)	(%SV)
Total Gage R&R	0.045625	0.27375	25.14
Repeatability <sup>.</sup>	0.033972	0.20383	18.72
Reproducibility	0.030455	0.18273	16.78
Part-To-Part	0.175629	1.05377	96.79
Total Variation	0.181458	1.08875	100.00

Number of Distinct Categories = 5

#### Gage R&R (Xbar/R) Report for Time (hours)



Reported by: Tolerance: Misc:



Procedure



#### Time (hours) by Official



Procedure \* Official Interaction



#### Gage R&R (ANOVA) Report for Time (hours)

Gage name: Date of study: Reported by: Tolerance: Misc:



### Summary of Results and Analysis

The following analysis was conducted following According to the Automobile Industry Action Group (AIAG) criteria:

If the Total Gage R&R contribution in the %Study Var column (% Tolerance, %Process) is:

- Less than 10% the measurement system is acceptable.
- Between 10% and 30% the measurement system is acceptable depending on the application, the cost of the measuring device, cost of repair, or other factors.
- Greater than 30% the measurement system is unacceptable and should be improved.

The total Gage R&R was 32.66%. Therefore the measurement system is unacceptable. Considerations may be made to improve the process, to lower the total Gage R&R to 10% or less.

#### Gage R&R Study - ANOVA Method

#### Two-Way ANOVA Table With Interaction

 Source
 DF
 SS
 MS
 F
 P

 PERSON
 9
 10.1251
 1.12501
 15.6594
 0.000

 OPERATOR
 2
 1.1085
 0.55425
 7.7148
 0.004

 PERSON \* OPERATOR
 18
 1.2932
 0.07184
 1.7416
 0.046

 Repeatability
 90
 3.7125
 0.04125
 7
 119
 16.2392

Alpha to remove interaction term = 0.25

#### Gage R&R

		<pre>%Contribution</pre>
Source	VarComp	(of VarComp)
Total Gage R&R	0.060958	40.99
Repeatability	0.041250	27.74
Reproducibility	0.019708	13.25
OPERATOR	0.012060	8.11
OPERATOR * PERSON	0.007648	5.14
Part-To-Part	0.087764	59.01
Total Variation	0.148722	100.00

		Study Var	Study Var
Source	StdDev (SD)	(6 * SD)	(% SV)
Total Gage R&R	0.246897	1.48138	64.02
Repeatability	0.203101	1.21861	52.67
Reproducibility	0.140386	0.84232	36.40
OPERATOR	0.109819	0.65891	28.48
OPERATOR * PERSON	0.087454	0.52472	22.68
Part-To-Part	0.296250	1.77750	76.82
Total Variation	0.385645	2.31387	100.00



#### Gage R&R (ANOVA) for MEASUREMENT

Gage name: Date of study: Reported by: Tolerance: Misc:



Welcome to Minitab press F1 for help

## Thank-You

# Questions?